




# WT389 groundplan

Review: B



| Ordernumber | Wind speed [m/s] | Current [A] | Review |
|-------------|------------------|-------------|--------|
| WT389-20-B  | 2...20           | 32*         | B      |
| WT389-30-B  | 3...30           | 64*         | B      |
| WT389-40-B  | 4...40           | 125*        | B      |
| WT389-50-B  | 5...50           | 200*        | B      |

\*estimated values

Specifications: (Customizing on request)

|                                          |          |           |
|------------------------------------------|----------|-----------|
| Working Section                          |          |           |
| Width                                    | [mm]     | 1200      |
| Height                                   | [mm]     | 1200      |
| Length                                   | [mm]     | 1650      |
| Air Data                                 |          |           |
| Contraction                              | [-]      | 2,78      |
| Turbulence Level<br>(at 50 % wind speed) | [%]      | 1         |
| Electrical Data                          |          |           |
| Supply Voltage                           | [V/50Hz] | 400       |
| Temperature                              |          |           |
| Operating Temp.                          | [°C]     | -10...+40 |